

**KASTURBA HEALTH SOCIETY'S  
MAHATMA GANDHI INSTITUTE OF MEDICAL SCIENCES,  
P.O. SEVAGRAM, WARDHA (DISTT.)  
MAHARASHTRA STATE - 442 102**

---

**Specifications**

**Anaesthesiology Department**

**Details: 1) Fully Automatic Oxygen Control Panel - Qty. 2 Nos.**

**Technical Specification:**

- 1) The Control Panel should be made to provide heavy duty and have a flow capacity of 2000 LPM or more at 50 to 60 PSI.
- 2) The fully automatic oxygen control panel should comply with HTM 02-01/NFPA99C/EN/DIN/ICO 7396-1 STANDARD.
- 3) All high pressure manifold regulators should contain no halogenated polymers.
- 4) All functional components should be enclosed in corrosion resistant robust material.
- 5) Fully automatic oxygen control panel the manifold assembly should provide two stages of pressure regulation. A single stage primary regulator one for each cylinder bank should be used to initially reduce cylinder pressure and two single stage pressure regulators should be provided in the control cabinet for final delivery pressure regulation. One delivery pressure regulator in service and one should be ready for service in a standby mode. The manifold control panel should be digital fully automatic type and switches from Bank in Use to Reserve bank without fluctuation in delivery supply line pressure. Changeover should be performed by electrically/Pneumatically operated valves contained in the control cabinet. In the event of an electrical power failure the valves should automatically open to provide an uninterrupted gas flow. The manifold should not require any manual resetting or adjustment after the replacements of the depleted cylinders. It should be US FDA/ European CE certified with 4 digit notified body number or American ETL/ American UL listed
- 6) Fully Automatic control panel should connect to our existing manifold.
- 7) Supplier should have installation base of one thousand bedded hospital.
- 8) Uptime Guarantee should be 100%.
- 9) Response time should be less than 24 hours
- 10) Time to rectification should be less than 48 hours.
- 11) If Time to rectification is more than 48 hours vendor should provide standby unit.
- 12) Quality assurance certificate to be provide from manufacturer.
- 13) Operating training provide to Gas plant operator
- 14) User and service manual to be provided.
- 15) Supplier should have the installation base at least 45 institution with 1000 bedded hospital.
- 16) Experience of providing after sales service of at least 10 years.

**Check List (to be submitted with Technical Bid)**

**The Vendor should submit following documents (duly attested) with Technical Bid**

(Please give Annexure No. for submitted Documents as mentioned below)

Name of Company: \_\_\_\_\_

Name of Contact Person: \_\_\_\_\_

Mobile No: \_\_\_\_\_ eMail ID: \_\_\_\_\_

Sr No	Name of Document	Document Submitted (Yes/No)	Annexure No	Reason for not submitting documents
1	Detailed technical specifications		I	
2	Authorized Distributor Certificate (Current Year's)		II	
3	User List (wherever Equipment / Instruments supplied & installed)		III	
4	Manufacturing Certificate of Manufacturer Company		IV	
5	Last 3 years Tax Assessment and Turn Over Certificates		V	
6	GST Registration Certificate		VI	
7	PAN Card		VII	
8	MSE Registration Certificate (if applicable)		VIII	
9	ISO Certification (ISO 9001 & ISO 13485)		IX	

**Signature with Company Seal**

**Kasturba Health Society's  
Mahatma Gandhi Institute of Medical Sciences  
Sevagram, Wardha (Dist), Maharashtra - 442 102**

**Terms and Conditions for the tenders**

**Any non compliance of the following will amount to disqualification of the tender**

1. Tender should be addressed to the **DEAN, Mahatma Gandhi Institute of Medical Sciences, Sevagram, (P.O.) Wardha (Dist), Maharashtra - 442 102**
2. Tender should be submitted by email at [centralstore@mgims.ac.in](mailto:centralstore@mgims.ac.in) on or before **30/04/2021 up to 12.00 Noon**. Tender received after the last date & time will not be accepted.
3. **The technical bid must contain the following (duly attested) along with Check List:**
  - a) Detailed technical specifications (Catalogue / illustrated literature) and options / other model.
  - b) A soft copy (CD) in Microsoft Word / Excel Format showing compliance with our given Technical Specifications (i.e. Compliance chart).
  - c) **Authorized dealer certificate (Current year's)** from the Principal of manufacturing company for supply and service
  - d) **List of users / institutional installations / Govt. supply orders / Reputed private hospitals and copy of Manufacturing License from the manufacturer.**
  - e) Last 3 years Tax Assessment and Turn Over Certificates
  - f) GST Registration Certificate
  - g) PAN Card copy
  - h) Country of Origin of the equipment
  - i) **If you are registered firm under Micro, Small Enterprises (MSEs), please submit a copy of the Registration Certificate.**
  - j) ISO Certification (ISO 9001 & ISO 13485)
4. **For Article / goods to be imported the prices should to be quoted in both**
  - a) **Foreign Currency - CIF Mumbai and also**
  - b) **INR - Delivery at Sevagram including all taxes and levies**

**Note:**

- a) FOR CIF prices, the duty, transportation from Mumbai port to Sevagram, insurance will be borne by the buyer.
- b) FOR INR prices-- no excise duty, GST, transportation, insurance will be paid by the buyer.
- c) MGIMS is holding a DSIR Certificate for Custom Duty concession and excise duty exemption.
- d) For ease of computation you may furnish the import duty structure of your equipment and classification HSN code no of custom tariff. Import License and IEC code should also be submitted.
- e) Kindly quote the prices strictly as per the scope of the supply written in the tender without offering your special features and accessories prices.
- f) You have to provide Performance Bank Guarantee covering the Warranty period
- g) Penalty clause for delay in supplying will be applicable.

**5. Payment terms:**

- a) For INR quote: Payment will be against successful installation of the equipment

b) No advances are given with Purchase Order (PO).

7. **Delivery:** Forwarding free on Road destination, i.e. door delivery basis
8. **Delivery period** from date of Purchase Order (PO) to be written clearly in days. Penalty clause can be levied on case to case basis. Installation of the equipment should be free and to the satisfaction of the buyer.
9. **Service and Guarantee conditions:**
  - **Equipment will be under guarantee as mentioned - Warranty should be 3 Years and 2 years CMC Free.** On case to case basis Performance Bank Guarantee may be sought.
  - **Mention the charges and terms of AMC & CMC (Labour + Part) applicable after expiry of warranty period.**
  - **Mention about the nearest facility/office where service personnel are available.**
  - **Mention about the facility for training of personnel safety and quality standards.**
10. **Dean M.G.I.M.S. reserves the right to reject all or any tender without assigning any reason.**